

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

AUG 29 2009

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130133

Name Bryan BARLOW

ID# 4548

Date 8/27/09

Agency Alaska State Troopers

Phone # 907-783-0972

Instrument Location Girdwood - AST

Alco S/N 75260

Target Value .079

High Pressure 700

Alco Test Value Average .078

1st Alco

.079

2nd Alco

Signature

Bryan Barlow

CRB
9/1/09

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (Do Not write in the area below)
- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
 - (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
 - (3) I am the Scientific Director of the State Breath Alcohol Program.
 - (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
 - (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
 - (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz

Scientific Director

State Breath Alcohol Program


Subscribed and sworn before me this _____ day of _____, 2009.

(Notary Seal Stamp)

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130133 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

[Empty box for taping the test strip]

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

AUGUST 27, 2009

OPERATOR'S NAME: **BARLOW BRYAN W**
OPERATOR'S NUMBER: 4548
SUBJECT'S LAST NAME: **CALIBRATION**
SUBJECT'S FIRST NAME/MI: **VERIFICATION OF**
O.L. #: 1234568
DEPT/AGENCY: 6040
CASE/REPORT: 09-12345
TEST TYPE: U
ALCO TARGET VALUE: .079
ALCO S/N: 75268

— BREATH ANALYSIS —

.079 ADJUSTED FOR 29.93 in
ALCO TARGET .079 22:23
BLANK TEST .000 22:24
INTERNAL STANDARD VERIFIED 22:24
ALCO TV 29.91 in .078 22:24
BLANK TEST .000 22:25
SUBJECT SAMPLE .000 22:26
BLANK TEST .000 22:27
ALCO TV 29.91 in .079 22:27
BLANK TEST .000 22:28

DIAGNOSTIC CHECK

[Empty box for taping the diagnostic check strip]

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130133

AUGUST 27, 2009
TIME 22:18

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 42c
BAROMETER: 29.91 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z \] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~ ¡ ¢